

PRESTO® W80

Cooling a 5 liters reactor from +50 °C to +20 °C

Objective

This case study tests the cooling power of PRESTO® W80 with a 5 l glass reactor. The PRESTO® W80 is connected to the reactor via two 1 m metal tubings. The PRESTO® W80 is programmed to cool down from +50 °C to +20 °C.

Environment

Room temperature +20 °C Humidity 45 %

Voltage 230 V / 50 Hz



Test Conditions

JULABO unit PRESTO® W80
Cooling power +20 °C 1.2 kW

0 °C 1.2 kW -20 °C 1.1 kW

Heating capacity 1.8 kW
Band limit without
Flow pressure 0.4 bar

Bath fluid Thermal HL 80

Reactor 5 l glass reactor (Rettberg)

filled with 5 I Thermal HL 80

Jacket volume 2.5 l

Control External (ICC)







Test Results

The PRESTO® W80 cooling process from +50 °C to +20 °C in 30 min without overshoot.





